



# Towards living cell sensors

## Simulating 'whispering gallery modes' in micro-resonators.

Dr Jonathan Hall: <http://drjonathanmmhallfrsa.wordpress.com>

**ARC Georgina Sweet Laureate group:** T. Monroe, S. Afshar, A. François, N. Riesen



## Sensing with light


- Light can **resonate** inside microscopic devices: spheres, disks or shells.
- **What for?** Resonators act as **detectors** of nearby macro-molecules, such as viruses, bacteria or DNA.
- **How?** Resonators of a certain size (diam: 5-30  $\mu\text{m}$ ) can support special resonant 'whispering gallery modes'.  
→ We can fabricate these resonators (e.g. polystyrene). Can we use **living cells**?



Illustration of a resonator. Polystyrene microspheres have been shown to **lase**.

## 'Whispering gallery modes'


- Electromagnetic waves at the **boundary** of a sphere or disk can be reflected around the surface.
- These resonant 'bound' whispering gallery modes correspond to the number of **surface nodes**, and **radial nodes**. They are also **narrow** ('high Q') and trackable.
- At the material/medium interface, an 'evanescent field' extends outward, which is **sensitive to the external environment**.



Whispering galleries

## Simulation methods

- To test the viability of a resonator cost-effectively, we have developed a simulation tool (based on 'FDTD').
- We choose the source type, wavelength, and resonator size, and measure the radiation through a flux region.
- Spheres, shells and odd-shape configurations are permitted, including inhomogeneous materials.




Collecting flux from a microsphere.

## Tools of the trade


- FDTD is good for 'higher index-contrast' scenarios, where *diameter*  $\div$  *wavelength* is not too large.
  - Analytic models (Mie Scattering/Shell-model) are good for 'lower index-contrast' scenarios, when *diameter*  $\div$  *wavelength* becomes large.
- BUT:
- If we have high index-contrast, and a large diameter compared to wavelength, modes are so narrow and closely-spaced we can't track them!

# Results

# Wavelength spectrum





## Wavelength spectrum






## Radiation distribution

- We can also **measure** how the power is distributed, e.g. as seen by a fibre.
- More concentrated modes (smaller angular distribution) are **less sensitive** to changes in large collection apertures.




Power distribution for four different modes (wavelengths).

## Fluorescent micro-shell simulation



## Fluorescent micro-shell simulation



## Plan for the future & wishlist


- **What we can do:**

We are mapping out resonator configurations **suitable for bio-sensing**. Realistic structural imperfections are incorporated.

- Cells that **match viable design solutions** will be sent for experiment.

**What we could use:**

- *Biology*: sourcing & sorting cells, genetic-engineering.
- *Medical*: mobilising device & physical properties of bio-markers.



Sensing technology of the future?